TC2MHR144-C

OPTO ENGINEERING

High resolution bi-telecentric lens for 1" detectors, magnification 0.089x, C-mount

SPECIFICATIONS

Magnification	(x)	0.089
Image circle Ø	(mm)	16.8
Object field of view 8		
with KAI-2020 14.8 mm diagonal w x h 11.84 x 8.88	(mm x mm)	133.5 x 100.1
with KAI-04050 16 mm diagonal w x h 12.8 x 9.6	(mm x mm)	144.3 x 108.2
with KAI-4022/4021 21.5 mm diagonal w x h 15.2 x 15.2	(mm x mm)	Ø=171.4
with KAI-08050 22.6 mm diagonal w x h 18.1 x 13.6(7)	(mm x mm)	Ø=153.3
Ontical specifications		







Optical specifications

Working distance (1)	(mm)	396
f/# (2)	· · · · · · ·	16
Telecentricity typical (max) (3)	(deg)	< 0.05 (0.10)
Distortion typical (max) (4)	(%)	< 0.05 (0.10)
Field depth (5)	(mm)	151
CTF@ 50 lp/mm	(%)	> 40

Mechanical specifications

Mount		С
Length (6)	(mm)	510.8
Diameter	(mm)	200
Mass	(g)	5895

NOTES

- 1. Working distance: distance between the front end of the mechanics and the object. Set this distance within +/- 3% of the nominal value for maximum resolution and minimum distortion.
- 2. Working F-number: the real F-number of a lens when used as a macro. Lenses with smaller apertures can be supplied on request.
- 3. Maximum slope of chief rays inside the lens: when converted to milliradians, it gives the maximum measurement error for any millimeter of object displacement. Typical (average production) values and maximum (guaranteed) values are listed.
- 4. Percent deviation of the real image compared to an ideal, undistorted image: typical (average production) values and maximum (guaranteed) values are listed.
- 5. At the borders of the field depth the image can be still used for measurement but, to get a perfectly sharp image, only half of the nominal field depth should be considered. Pixel size used for calculation is 5 um.
- 6. Measured from the front end of the mechanics to the camera flange.
- 7. With KAI-08050 (22.6 mm diagonal) detectors, the FOV of TC4MHR yyy lenses may show some vignetting at the image corners, as these lenses are optimized for 1.2" detectors (21.5 mm diagonal).
- 8. For the fields with the indication "Ø =", the image of a circular object of such diameter is fully inscribed into the detector.

COMPATIBLE PRODUCTS



LTCLHP144-G Telecentric HP illuminator, beam diameter 180 mm, green



CMHO144 Clamping mechanics for TCxx130, TCxx144 lenses and LTCLHP144-X illuminators



LTRN144NW Ring LED illuminator, white