TC 23 009

Bi-telecentric lens for 2/3" detectors, magnification 1.000 x, C-mount

SPECIFICATIONS

Magnification

Image circle Ø	(mm)	11.0
Object field of view (8)		
with 1/3" detector (4.8 x 3.6 mm)	(mm x mm)	4.80 x 3.60
with 1/2.5" detector (5.70 x 4.28 mm)	(mm x mm)	5.70 x 4.28
with 1/2" detector (6.4 x 4.8 mm)	(mm x mm)	6.40 x 4.80
with 1/1.8" detector (7.13 x 5.37 mm) (7)	(mm x mm)	7.13 x 5.37
with 2/3" - 5 Mpx detector (8.45 x 7.07 mm)	(mm x mm)	8.44 x 7.06

(x)


1.000



Working distance (1)	(mm)	62.2
f/# (2)		11
Telecentricity typical (max) (3)	(deg)	< 0.08 (0.10)
Distortion typical (max) (4)	(%)	< 0.04 (0.08)
Field depth (5)	(mm)	0.9
CTF @ 70 lp/mm	(%)	> 25

Dimensions		
Mount		С
Length (6)	(mm)	65.0
Diameter	(mm)	28
Mass	(g)	160









NOTES

- 1. Working Distance: distance between the front end of the mechanics and the object. Set this distance within +/- 3% of the nominal value for maximum resolution and minimum distortion.
- $2. \ \ Working \ F-number: the \ real \ F-number \ of \ a \ lens \ when \ used \ as \ a \ macro. \ Lenses \ with \ smaller \ apertures$ can be supplied on request.
- ${\it 3. }\ Maximum\ slope\ of\ chief\ rays\ inside\ the\ lens:\ when\ converted\ to\ millirad,\ it\ gives\ the\ maximum$ measurement error for any millimeter of object displacement. Typical (average production) values and maximum (guaranteed) values are listed.
- 4. Percent deviation of the real image compared to an ideal, undistorted image: typical (average production) values and maximum (guaranteed) values are listed.
- 5. At the borders of the field depth the image can be still used for measurement but, to get a very sharp image, only half of the nominal field depth should be considered. Pixel size used for calculation is 5.5
- 6. Measured from the front end of the mechanics to the camera flange.
- 7. With 1/1.8" (9 mm diagonal) detectors, the FOV of TC12 yyy lenses may show some vignetting at the image corners, as these lenses are optimized for 1/2" detectors (8 mm diagonal).
- 8. For the fields with the indication "Ø =", the image of a circular object of such diameter is fully inscribed into the detector.

COMPATIBLE PRODUCTS



LTCLHP023-G Telecentric HP illuminator, beam diameter 16 mm, green



CMHO 023 Clamping mechanics for TCxx004, 007, 009, 012 and LTCLHP023-X



LTRN 023 NW Ring LED illuminator, white