# TC4MHR240-F



High resolution bi-telecentric lens for 1.2" detectors, magnification 0.073x, F-mount

#### **SPECIFICATIONS**

Magnification

Image circle Ø	(mm)	21.1
Object field of view 8		
with KAI-2020 14.8 mm diagonal w x h 11.84 x 8.88	(mm x mm)	161.7 x 121.3
with KAI-04050 16 mm diagonal w x h 12.8 x 9.6	(mm x mm)	174.9 x 131.1
with KAI-4022/4021 21.5 mm diagonal w x h 15.2 x 15.2	(mm x mm)	207.7 x 207.7
with KAI-08050 22.6 mm diagonal w x h 18.1 x 13.6(7)	(mm x mm)	247.3 x 185.8

(x)

0.073







### Optical specifications

Working distance (1)	(mm)	492.86
f/# (2)		16
Telecentricity typical (max) (3)	(deg)	< 0.05 (0.10)
Distortion typical (max) (4)	(%)	< 0.05 (0.10)
Field depth (5)	(mm)	221.5
CTF@ 50 lp/mm	(%)	> 30

#### Mechanical specifications

·		
Mount		F
Length (6)	(mm)	798.3
Diameter	(mm)	322
Mass	(g)	18474

## NOTES

- 1. Working distance: distance between the front end of the mechanics and the object. Set this distance within +/- 3% of the nominal value for maximum resolution and minimum distortion.
- 2. Working F-number: the real F-number of a lens when used as a macro. Lenses with smaller apertures
- 3. Maximum slope of chief rays inside the lens: when converted to milliradians, it gives the maximum measurement error for any millimeter of object displacement. Typical (average production) values and maximum (guaranteed) values are listed.
- 4. Percent deviation of the real image compared to an ideal, undistorted image: typical (average production) values and maximum (guaranteed) values are listed.
- 5. At the borders of the field depth the image can be still used for measurement but, to get a perfectly sharp image, only half of the nominal field depth should be considered. Pixel size used for calculation is  $5\,\mu m$ .
- $\,$  6. Measured from the front end of the mechanics to the camera flange.
- 7. With KAI-08050 (22.6 mm diagonal) detectors, the FOV of TC4MHR yyy lenses may show some vignetting at the image corners, as these lenses are optimized for 1.2" detectors (21.5 mm diagonal).
- 8. For the fields with the indication "Ø =", the image of a circular object of such diameter is fully inscribed into the detector.

# COMPATIBLE PRODUCTS



LTCLHP240-G Telecentric HP illuminator, beam diameter 300 mm, green